

**TDK-Lambda Americas**

High Power

405 Essex Road Neptune, NJ 07753

TEL. +1 732 922 9300 ~ FAX +1 732 922 9334

Email: rita.muraco@us.tdk-lambda.com

Internet <http://www.us.tdk-lambda.com/hp>**Repair Quotation No. 2018049**

Date: 02/09/18

To: Duane Thompson
Company: Monterey Bay Aquarium
Address: Moss Landing, CA
Telephone: 831-775-2003
Email: drthom@mbari.org

From: Rita Muraco
Phone No. 732-922-9300 Ext. 325

Dear Duane,

Please find below estimate for **RMA009543** to further repair your unit. Please be advice that this is a **preliminary** estimate and costs are subject to change.

UNIT FOR REPAIR	SERIAL NO#	PRELIMINARY REPORT	PRELIMINARY ESTIMATE
GEN 40-60-LAN-1P230	LOC-633A277-0002	Upon evaluation, the technician confirmed unit “ will not power on” The PFC board has 4 open inrush resistors and 3 blown off runs. The PCA connection DC/DC to PFC has a burnt run. This failure was caused due to the environment is not covered under warranty.	\$1,056.00

This is a **preliminary** estimate, the charges stated on this quotation, covers only the above repair and parts. If for any reason failures are found upon testing, you will be notified before any other repairs are made.

Estimated Lead Time of Repair

4-6 Weeks Upon Approval**NOTES:**

Purchase Orders to be made out to: TDK-Lambda Americas Inc. 405 Essex Road, Neptune, NJ 07753

TERMS: Net 30 Days pending Credit Check or Credit Card required for repairs under \$750.00

F.O.B. Neptune, NJ

Freight, Handling and Insurance: for Customer's Account and Responsibility

TAXES: Prices shown do not include any federal, state or local taxes, if applicable.

SHIPMENTS: Via least expensive surface transportation

Should you need further assistance or have any questions, please do not hesitate to contact me.

Best regards,