

Paroscientific, Inc.

Pressure Instrument Configuration

SN: 87760 Part Number: 1789-003-0 Model: 8CB4000-I-257 Port: Oil Filled

Calibration Date: 26-May-21 Report No: 24442 Technician: RM

Pressure Range: 0 to 4000 meters Temperature Range: -2 to 40 deg C

Customer: MBARI Accounts Payable

Report Date: 26-May-21

Address : 7700 Sandholdt Road

Sales Order: 37749

Moss Landing, CA 95039-9644 USA

S/R Number : 12396

Configuration		Calibration Coefficients	
BL: 0	PT: N	U0: 5.845663	µsec
BR: 9600	QD: -	Y1: -3893.683	deg C / µsec
DD: -	QO: -	Y2: -10415.54	deg C / µsec ²
DL: 1	SL: 0	Y3: 0	deg C / µsec ³
DM: 0	SN: 87760	C1: -25399.16	psi
DO: 0	ST: 10	C2: -147.0909	psi / µsec
DP: 6	SU: 0	C3: 85063.95	psi / µsec ²
ID: 01	TI: 30	D1: 0.0398335	
IM: 0	TR: 40	D2: 0.0000000	
LL: .0000000	TU: 0	T1: 30.15023	µsec
LH: 2250.000	UF: 1.000000	T2: 0.898331	µsec / µsec
MC: -	UL:	T3: 59.70873	µsec / µsec ²
MD: 0	UM: ŸŸŸŸ	T4: 170.8274	µsec / µsec ³
MN: 8CB4000-I-257	UN: 3	T5: 0.000	µsec / µsec ⁴
OP: 420.5802	US: 1	TC: 0.6781680	
PF: 413.6855	VR: R5.10	PA: -0.0616203	
PI: 30	ZI: 0	PM: 0.9998950	
PL: 7200.000	ZS: 0		
PO: 0	ZL: 0		
PR: 10	ZV: .0000000000		
PS: 0			

Paroscientific, Inc.

(Data As Left)

Prepared by

4500 148th Ave. N.E. Redmond, WA 98052

Phone: (425)883-8700 Fax: (425)867-5407

Web: <http://www.paroscientific.com>

Email: support@paroscientific.com



Technology