



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0448
CALIBRATION DATE: 28-Oct-23

SBE 49 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 8.027697e-004
a1 = 2.944723e-004
a2 = -3.383895e-006
a3 = 2.410940e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	719012.869	1.0001	0.0001
4.4999	635562.033	4.4998	-0.0001
15.0000	424417.393	15.0001	0.0001
18.5000	366710.172	18.5000	-0.0000
24.0000	287636.393	23.9999	-0.0001
29.0000	226861.711	29.0000	0.0000
32.5000	189899.877	32.5000	0.0000

n = Instrument Output (counts)

$MV = (n - 524288) / 1.6e+007$

$R = (MV * 2.295e+010 + 9.216e+008) / (6.144e+004 - MV * 5.3e+005)$

Temperature ITS-90 (°C) = $1 / \{a0 + a1[\ln(R)] + a2[\ln^2(R)] + a3[\ln^3(R)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

