

ASIMET WND

1. Manually-configured instrument settings not handled by service

None.

2. Device metadata returned in SIAM datastream

```
WND01
WND358
VOSWND53 v3.5
2.4576 Mhz
22JUL05
2006/03/17 15:13:30
No PCMCIA card installed
```

```
MODADR: WND01
MODMFG: STARENG.
MODMOD: VOSWND
MODSER: WND358
MODDAT: 29JUL03
SENMFGE: USD/PNAV
SENMOD: A2-S-B/-
SENER: 9241/87
SENDAT: -
SFTMFGE: WHOI/GEOFF
SFTNAM: PICWND
SFTREV: V1.5
SFTDAT: -
CALFAC: WHOI
CALPER: J.C.SMITH
CALDAT: 22JUL05
DATFRM: -
DATDES: -
DATUNI: -
RAWFRM: -
RAWDES: -
RAWUNI: -
```

3. Instrument-specific service properties

```
/** Must be LWR, SWR, WND, HRH, or BPR */
String sensorType;
```

```
/** Instrument module address - "01", "02", etc. */
String moduleAddress;
```

4. Initial service property values for this deployment

```
powerPolicy=ALWAYS
commPowerPolicy=SAMPLING
currentLimitMa=2000
nominalCurrentMa=1000
```

```
powerOnDelaySec=0
sampleSchedule=R */0 */10 */0 */0 * * * * GMT *
maxSampleTries=3
sampleTimeoutMsec=3000
startDelayMsec=4000

sensorType=WND
moduleAddress=01
```

5. Results of acceptance test procedures

OK

6. Known-service bugs/deficiencies

None