

Data Report for Device 1534_1482 from Thu Mar 16 16:46:44 UTC 2006 to Fri Mar 17 16:46:44 UTC 2006 sorted by SIAM timestamp

SIAM Timestamp	SIAM Timestamp converted to time/date (PST)	Packet Sequence Number	String Representation of Data Buffer
1142533715449	2006-03-16 10:28:35 -0800	275	<pre> <DEV_MD> SBE37-SM-PUCK V 1.0-01 SERIAL NO. 2765 01 Jan 1996 06:28:31 logging not started sample interval = 60 seconds samplenumber = 0, free = 183202 do not transmit real-time data do not output salinity with each sample do not output sound velocity with each sample store time with each sample number of samples to average = 4 serial sync mode disabled wait time after serial sync sampling = 30 seconds PUCK mode enabled internal pump not installed temperature = 21.66 deg C SBE37-SM-PUCK V 1.0-01 2765 temperature: 12-dec-02 -4.215792e-05 2.839508e-04 -3.000133e-06 1.696669e-07 conductivity: 12-dec-02 -1.051158e+00 1.466496e-01 -9.150801e-05 3.157882e-05 -9.570000e-08 3.250000e-06 0.000000e+00 pressure S/N 2598, range = 160 psia: 09-dec-02 5.945600e-02 7.603297e-03 -7.831820e-10 5.468956e+01 1.021594e+00 -2.910683e-02 2.494063e+01 9.250000e-04 0.000000e+00 0.000000e+00 rtc: 12-dec-02 9.999726e-01 1.708880e-06 -3.062022e-08 </DEV_MD> <SRVC_MD> powerOnDelaySec=0 instrumentName=UNKNOWN nominalCurrentMa=1000 className=org.mbari.isi.interfaces.InstrumentServiceAttributes serviceName=Seabird-37SM s/n 37SM30960-2765 parentID=1482 currentLimitMa=5000 sampleSchedule=R */0 */10 */0 */0 * * * * GMT * extendedVersion=UNK frameworkVersion=\$Name: march14 2006 \$ UUID=5832530d-ad76-11da-bb39-00306e389969 powerPolicy=ALWAYS timeSynch=false maxSampleTries=3 isiID=1534 sampleTimeoutMsec=15000 commPowerPolicy=SAMPLING serviceStatus=2 startDelayMsec=2000 </SRVC_MD> <DOC_MD> unable to access XML file </DOC_MD> </pre>
1142533723456	2006-03-16 10:28:43 -0800	276	21.6604, 0.00005, 0.089, 01 Jan 1996, 06:28:36
1142534324367	2006-03-16 10:38:44 -0800	277	21.7126, 0.00005, 0.091, 01 Jan 1996, 06:38:36
1142534924358	2006-03-16 10:48:44 -0800	278	21.8011, 0.00005, 0.091, 01 Jan 1996, 06:48:38
1142535526129	2006-03-16 10:58:46 -0800	279	21.9052, 0.00005, 0.093, 01 Jan 1996, 06:58:38
1142536124358	2006-03-16 11:08:44 -0800	280	21.9542, 0.00005, 0.095, 01 Jan 1996, 07:08:36
1142536724390	2006-03-16 11:18:44 -0800	281	22.0916, 0.00005, 0.086, 01 Jan 1996, 07:18:37
1142537324258	2006-03-16 11:28:44 -0800	282	22.1913, 0.00005, 0.089, 01 Jan 1996, 07:28:36
1142537924307	2006-03-16 11:38:44 -0800	283	22.3038, 0.00005, 0.091, 01 Jan 1996, 07:38:38
1142538524394	2006-03-16 11:48:44 -0800	284	22.4123, 0.00005, 0.084, 01 Jan 1996, 07:48:38
1142539124378	2006-03-16 11:58:44 -0800	285	22.4980, 0.00005, 0.086, 01 Jan 1996, 07:58:38
1142539724417	2006-03-16 12:08:44 -0800	286	22.6393, 0.00005, 0.088, 01 Jan 1996, 08:08:38
1142540325535	2006-03-16 12:18:45 -0800	287	22.7520, 0.00005, 0.080, 01 Jan 1996, 08:18:38
1142540924291	2006-03-16 12:28:44 -0800	288	22.9163, 0.00005, 0.083, 01 Jan 1996, 08:28:38
1142541524303	2006-03-16 12:38:44 -0800	289	23.0901, 0.00005, 0.087, 01 Jan 1996, 08:38:38

1142542124351	2006-03-16 12:48:44 -0800	290	23.2510, 0.00005, 0.080, 01 Jan 1996, 08:48:38
1142542724291	2006-03-16 12:58:44 -0800	291	23.4915, 0.00005, 0.085, 01 Jan 1996, 08:58:37
1142543324330	2006-03-16 13:08:44 -0800	292	23.4484, 0.00005, 0.076, 01 Jan 1996, 09:08:37
1142543924341	2006-03-16 13:18:44 -0800	293	23.5771, 0.00005, 0.078, 01 Jan 1996, 09:18:37
1142544524348	2006-03-16 13:28:44 -0800	294	23.6524, 0.00005, 0.071, 01 Jan 1996, 09:28:37
1142545124378	2006-03-16 13:38:44 -0800	295	23.5857, 0.00005, 0.071, 01 Jan 1996, 09:38:37
1142545724336	2006-03-16 13:48:44 -0800	296	23.7982, 0.00005, 0.073, 01 Jan 1996, 09:48:38
1142546326054	2006-03-16 13:58:46 -0800	297	23.7011, 0.00005, 0.073, 01 Jan 1996, 09:58:38
1142546924296	2006-03-16 14:08:44 -0800	298	23.7641, 0.00005, 0.063, 01 Jan 1996, 10:08:36
1142547524313	2006-03-16 14:18:44 -0800	299	23.7339, 0.00005, 0.062, 01 Jan 1996, 10:18:36
1142548124544	2006-03-16 14:28:44 -0800	300	23.7917, 0.00005, 0.062, 01 Jan 1996, 10:28:38
1142548724383	2006-03-16 14:38:44 -0800	301	23.8547, 0.00005, 0.063, 01 Jan 1996, 10:38:38
1142549324264	2006-03-16 14:48:44 -0800	302	23.7176, 0.00005, 0.063, 01 Jan 1996, 10:48:38
1142549924330	2006-03-16 14:58:44 -0800	303	23.8094, 0.00005, 0.063, 01 Jan 1996, 10:58:38
1142550524326	2006-03-16 15:08:44 -0800	304	23.9574, 0.00005, 0.063, 01 Jan 1996, 11:08:38
1142551124322	2006-03-16 15:18:44 -0800	305	23.8081, 0.00005, 0.063, 01 Jan 1996, 11:18:38
1142551724341	2006-03-16 15:28:44 -0800	306	23.6317, 0.00005, 0.061, 01 Jan 1996, 11:28:38
1142552324436	2006-03-16 15:38:44 -0800	307	23.6872, 0.00005, 0.059, 01 Jan 1996, 11:38:38
1142552924377	2006-03-16 15:48:44 -0800	308	23.7151, 0.00005, 0.058, 01 Jan 1996, 11:48:38
1142553524313	2006-03-16 15:58:44 -0800	309	23.7481, 0.00005, 0.057, 01 Jan 1996, 11:58:37
1142554124341	2006-03-16 16:08:44 -0800	310	23.7873, 0.00005, 0.057, 01 Jan 1996, 12:08:37
1142554724513	2006-03-16 16:18:44 -0800	311	23.8120, 0.00005, 0.057, 01 Jan 1996, 12:18:38
1142555324323	2006-03-16 16:28:44 -0800	312	23.8354, 0.00005, 0.056, 01 Jan 1996, 12:28:37
1142555924354	2006-03-16 16:38:44 -0800	313	23.8612, 0.00005, 0.057, 01 Jan 1996, 12:38:37
1142556524374	2006-03-16 16:48:44 -0800	314	23.8755, 0.00005, 0.057, 01 Jan 1996, 12:48:37
1142557124343	2006-03-16 16:58:44 -0800	315	23.8872, 0.00005, 0.058, 01 Jan 1996, 12:58:37
1142557724263	2006-03-16 17:08:44 -0800	316	23.8966, 0.00005, 0.058, 01 Jan 1996, 13:08:37
1142558324303	2006-03-16 17:18:44 -0800	317	23.8653, 0.00005, 0.058, 01 Jan 1996, 13:18:38
1142558924387	2006-03-16 17:28:44 -0800	318	23.8521, 0.00005, 0.057, 01 Jan 1996, 13:28:38
1142559525692	2006-03-16 17:38:45 -0800	319	23.5917, 0.00005, 0.054, 01 Jan 1996, 13:38:38
1142560124331	2006-03-16 17:48:44 -0800	320	23.5486, 0.00005, 0.051, 01 Jan 1996, 13:48:38
1142560724548	2006-03-16 17:58:44 -0800	321	23.4851, 0.00005, 0.050, 01 Jan 1996, 13:58:38
1142561324308	2006-03-16 18:08:44 -0800	322	23.3879, 0.00005, 0.049, 01 Jan 1996, 14:08:38
1142561924542	2006-03-16 18:18:44 -0800	323	23.3047, 0.00005, 0.049, 01 Jan 1996, 14:18:38
1142562524318	2006-03-16 18:28:44 -0800	324	23.2672, 0.00005, 0.048, 01 Jan 1996, 14:28:38
1142563124302	2006-03-16 18:38:44 -0800	325	23.1897, 0.00005, 0.048, 01 Jan 1996, 14:38:38
1142563724327	2006-03-16 18:48:44 -0800	326	23.1571, 0.00005, 0.047, 01 Jan 1996, 14:48:38
1142563889550	2006-03-16 18:51:29 -0800	327	<DEV_MD> SBE37-SM-PUCK V 1.0-01 SERIAL NO. 2765 01 Jan 1996 02:51:25 logging not started sample interval = 60 seconds sampler number = 0, free = 183202 do not transmit real-time data do not output salinity with each sample do not output sound velocity with each sample store time with each sample number of samples to average = 4 serial sync mode disabled wait time after serial sync sampling = 30 seconds PUCK mode enabled internal pump not installed temperature = 23.15 deg C SBE37-SM-PUCK V 1.0-01 2765 temperature: 12-dec-02 -4.215792e-05 2.839508e-04 -3.000133e-06 1.696669e-07 conductivity: 12-dec-02 -1.051158e+00 1.466496e-01 -9.150801e-05 3.157882e-05 -9.570000e-08 3.250000e-06 0.000000e+00 pressure S/N 2598, range = 160 psia: 09-dec-02 5.945600e-02 7.603297e-03 -7.831820e-10 5.468956e+01 1.021594e+00 -2.910683e-02 2.494063e+01 9.250000e-04 0.000000e+00 0.000000e+00 rtc: 12-dec-02 9.999726e-01 1.708880e-06 -3.062022e-08 </DEV_MD>

			<pre><SRVC_MD> powerOnDelaySec=0 instrumentName=UNKNOWN nominalCurrentMa=1000 className=org.mbari.isi.interfaces.InstrumentServiceAttributes serviceName=Seabird-37SM s/n 37SM30960-2765 parentID=1482 currentLimitMa=5000 sampleSchedule=R */0 */10 */0 */0 * * * * GMT * extendedVersion=UNK frameworkVersion=\$Name: march14_2006 \$ UUID=5832530d-ad76-11da-bb39-00306e389969 powerPolicy=ALWAYS timeSynch=false maxSampleTries=3 isiID=1534 sampleTimeoutMsec=15000 commPowerPolicy=SAMPLING serviceStatus=2 startDelayMsec=2000 </SRVC_MD> <DOC_MD> unable to access XML file </DOC_MD></pre>
1142563897938	2006-03-16 18:51:37 -0800	328	23.1490, 0.00005, 0.047, 01 Jan 1996, 02:51:30
1142564500625	2006-03-16 19:01:40 -0800	329	23.0767, 0.00005, 0.047, 01 Jan 1996, 03:01:31
1142565098640	2006-03-16 19:11:38 -0800	330	23.0590, 0.00005, 0.046, 01 Jan 1996, 03:11:29
1142565698711	2006-03-16 19:21:38 -0800	331	23.0365, 0.00005, 0.046, 01 Jan 1996, 03:21:29
1142566298656	2006-03-16 19:31:38 -0800	332	22.9649, 0.00005, 0.046, 01 Jan 1996, 03:31:29
1142566898648	2006-03-16 19:41:38 -0800	333	22.9127, 0.00005, 0.045, 01 Jan 1996, 03:41:29
1142567794541	2006-03-16 19:56:34 -0800	334	<pre><DEV_MD> SBE37-SM-PUCK V 1.0-01 SERIAL NO. 2765 01 Jan 1996 03:56:30 logging not started sample interval = 60 seconds samplenumber = 0, free = 183202 do not transmit real-time data do not output salinity with each sample do not output sound velocity with each sample store time with each sample number of samples to average = 4 serial sync mode disabled wait time after serial sync sampling = 30 seconds PUCK mode enabled internal pump not installed temperature = 22.90 deg C SBE37-SM-PUCK V 1.0-01 2765 temperature: 12-dec-02 -4.215792e-05 2.839508e-04 -3.000133e-06 1.696669e-07 conductivity: 12-dec-02 -1.051158e+00 1.466496e-01 -9.150801e-05 3.157882e-05 -9.570000e-08 3.250000e-06 0.000000e+00 pressure S/N 2598, range = 160 psia: 09-dec-02 5.945600e-02 7.603297e-03 -7.831820e-10 5.468956e+01 1.021594e+00 -2.910683e-02 2.494063e+01 9.250000e-04 0.000000e+00 0.000000e+00 rtc: 12-dec-02 9.999726e-01 1.708880e-06 -3.062022e-08 </DEV_MD> <SRVC_MD> powerOnDelaySec=0 instrumentName=UNKNOWN nominalCurrentMa=1000 className=org.mbari.isi.interfaces.InstrumentServiceAttributes serviceName=Seabird-37SM s/n 37SM30960-2765 parentID=1482 currentLimitMa=5000 sampleSchedule=R */0 */10 */0 */0 * * * * GMT * extendedVersion=UNK frameworkVersion=\$Name: march14_2006 \$ UUID=5832530d-ad76-11da-bb39-00306e389969 powerPolicy=ALWAYS timeSynch=false maxSampleTries=3 isiID=1534 sampleTimeoutMsec=15000 commPowerPolicy=SAMPLING serviceStatus=2</pre>

			startDelayMsec=2000 </SRVC_MD> <DOC_MD> unable to access XML file </DOC_MD>
1142567801961	2006-03-16 19:56:41 -0800	335	22.9013, 0.00005, 0.045, 01 Jan 1996, 03:56:34
1142568607910	2006-03-16 20:10:07 -0800	336	<DEV_MD> SBE37-SM-PUCK V 1.0-01 SERIAL NO. 2765 01 Jan 1996 04:10:03 logging not started sample interval = 60 seconds samplenummer = 0, free = 183202 do not transmit real-time data do not output salinity with each sample do not output sound velocity with each sample store time with each sample number of samples to average = 4 serial sync mode disabled wait time after serial sync sampling = 30 seconds PUCK mode enabled internal pump not installed temperature = 22.80 deg C SBE37-SM-PUCK V 1.0-01 2765 temperature: 12-dec-02 -4.215792e-05 2.839508e-04 -3.000133e-06 1.696669e-07 conductivity: 12-dec-02 -1.051158e+00 1.466496e-01 -9.150801e-05 3.157882e-05 -9.570000e-08 3.250000e-06 0.000000e+00 pressure S/N 2598, range = 160 psia: 09-dec-02 5.945600e-02 7.603297e-03 -7.831820e-10 5.468956e+01 1.021594e+00 -2.910683e-02 2.494063e+01 9.250000e-04 0.000000e+00 0.000000e+00 rtc: 12-dec-02 9.999726e-01 1.708880e-06 -3.062022e-08 </DEV_MD> <SRVC_MD> powerOnDelaySec=0 instrumentName=UNKNOWN nominalCurrentMa=1000 className=org.mbari.isi.interfaces.InstrumentServiceAttributes serviceName=Seabird-37SM s/n 37SM30960-2765 parentID=1482 currentLimitMa=5000 sampleSchedule=R */0 */10 */0 */0 * * * * GMT * extendedVersion=UNK frameworkVersion=\$Name: march14_2006 \$ UUID=5832530d-ad76-11da-bb39-00306e389969 powerPolicy=ALWAYS timeSynch=false maxSampleTries=3 isiID=1534 sampleTimeoutMsec=15000 commPowerPolicy=SAMPLING serviceStatus=2 startDelayMsec=2000 </SRVC_MD> <DOC_MD> unable to access XML file </DOC_MD>
1142568615436	2006-03-16 20:10:15 -0800	337	22.8068, 0.00005, 0.045, 01 Jan 1996, 04:10:07
1142569217183	2006-03-16 20:20:17 -0800	338	22.8109, 0.00005, 0.046, 01 Jan 1996, 04:20:09
1142569817174	2006-03-16 20:30:17 -0800	339	22.7289, 0.00005, 0.047, 01 Jan 1996, 04:30:09
1142570418940	2006-03-16 20:40:18 -0800	340	22.6938, 0.00005, 0.046, 01 Jan 1996, 04:40:09
1142571017333	2006-03-16 20:50:17 -0800	341	22.6989, 0.00005, 0.045, 01 Jan 1996, 04:50:10
1142571617254	2006-03-16 21:00:17 -0800	342	22.6606, 0.00005, 0.045, 01 Jan 1996, 05:00:11
1142572217317	2006-03-16 21:10:17 -0800	343	22.6703, 0.00005, 0.044, 01 Jan 1996, 05:10:11
1142572817381	2006-03-16 21:20:17 -0800	344	22.5882, 0.00005, 0.044, 01 Jan 1996, 05:20:11
1142573417193	2006-03-16 21:30:17 -0800	345	22.5865, 0.00005, 0.043, 01 Jan 1996, 05:30:11
1142574017278	2006-03-16 21:40:17 -0800	346	22.5728, 0.00005, 0.043, 01 Jan 1996, 05:40:10
1142574617273	2006-03-16 21:50:17 -0800	347	22.5595, 0.00005, 0.042, 01 Jan 1996, 05:50:10
1142575220632	2006-03-16 22:00:20 -0800	348	22.5403, 0.00005, 0.042, 01 Jan 1996, 06:00:11
1142575817241	2006-03-16 22:10:17 -0800	349	22.5067, 0.00005, 0.042, 01 Jan 1996, 06:10:11
1142576417554	2006-03-16 22:20:17 -0800	350	22.4802, 0.00005, 0.041, 01 Jan 1996, 06:20:11
1142577017292	2006-03-16 22:30:17 -0800	351	22.4423, 0.00005, 0.041, 01 Jan 1996, 06:30:10

1142577617258	2006-03-16 22:40:17 -0800	352	22.4379, 0.00005, 0.040, 01 Jan 1996, 06:40:11
1142578217453	2006-03-16 22:50:17 -0800	353	22.4020, 0.00005, 0.040, 01 Jan 1996, 06:50:11
1142578817348	2006-03-16 23:00:17 -0800	354	22.3919, 0.00005, 0.039, 01 Jan 1996, 07:00:10
1142579417392	2006-03-16 23:10:17 -0800	355	22.3392, 0.00005, 0.039, 01 Jan 1996, 07:10:10
1142580017474	2006-03-16 23:20:17 -0800	356	22.2953, 0.00005, 0.039, 01 Jan 1996, 07:20:11
1142580617348	2006-03-16 23:30:17 -0800	357	22.2280, 0.00005, 0.038, 01 Jan 1996, 07:30:10
1142581217278	2006-03-16 23:40:17 -0800	358	22.1628, 0.00005, 0.037, 01 Jan 1996, 07:40:10
1142581820601	2006-03-16 23:50:20 -0800	359	22.0831, 0.00005, 0.035, 01 Jan 1996, 07:50:10
1142582417408	2006-03-17 00:00:17 -0800	360	22.0473, 0.00005, 0.033, 01 Jan 1996, 08:00:10
1142583017353	2006-03-17 00:10:17 -0800	361	21.9336, 0.00005, 0.031, 01 Jan 1996, 08:10:11
1142583617368	2006-03-17 00:20:17 -0800	362	21.9174, 0.00005, 0.030, 01 Jan 1996, 08:20:11
1142584219251	2006-03-17 00:30:19 -0800	363	21.8710, 0.00005, 0.039, 01 Jan 1996, 08:30:10
1142584817339	2006-03-17 00:40:17 -0800	364	21.8122, 0.00005, 0.039, 01 Jan 1996, 08:40:10
1142585417426	2006-03-17 00:50:17 -0800	365	21.7383, 0.00005, 0.039, 01 Jan 1996, 08:50:10
1142586017318	2006-03-17 01:00:17 -0800	366	21.6453, 0.00005, 0.038, 01 Jan 1996, 09:00:10
1142586617326	2006-03-17 01:10:17 -0800	367	21.5853, 0.00005, 0.038, 01 Jan 1996, 09:10:10
1142587217443	2006-03-17 01:20:17 -0800	368	21.5223, 0.00005, 0.038, 01 Jan 1996, 09:20:10
1142587817168	2006-03-17 01:30:17 -0800	369	21.4565, 0.00005, 0.039, 01 Jan 1996, 09:30:11
1142588417278	2006-03-17 01:40:17 -0800	370	21.3957, 0.00005, 0.038, 01 Jan 1996, 09:40:11
1142589017283	2006-03-17 01:50:17 -0800	371	21.3264, 0.00005, 0.038, 01 Jan 1996, 09:50:11
1142589617313	2006-03-17 02:00:17 -0800	372	21.2731, 0.00005, 0.035, 01 Jan 1996, 10:00:09
1142590217257	2006-03-17 02:10:17 -0800	373	21.2063, 0.00005, 0.033, 01 Jan 1996, 10:10:10
1142590817305	2006-03-17 02:20:17 -0800	374	21.1706, 0.00005, 0.031, 01 Jan 1996, 10:20:09
1142591417682	2006-03-17 02:30:17 -0800	375	21.1099, 0.00005, 0.030, 01 Jan 1996, 10:30:11
1142592017658	2006-03-17 02:40:17 -0800	376	21.0908, 0.00005, 0.031, 01 Jan 1996, 10:40:11
1142592617312	2006-03-17 02:50:17 -0800	377	21.0197, 0.00005, 0.029, 01 Jan 1996, 10:50:10
1142593217223	2006-03-17 03:00:17 -0800	378	20.9640, 0.00005, 0.027, 01 Jan 1996, 11:00:10
1142593817372	2006-03-17 03:10:17 -0800	379	20.9269, 0.00005, 0.029, 01 Jan 1996, 11:10:10
1142594417433	2006-03-17 03:20:17 -0800	380	20.8981, 0.00005, 0.027, 01 Jan 1996, 11:20:10
1142595017213	2006-03-17 03:30:17 -0800	381	20.8325, 0.00005, 0.026, 01 Jan 1996, 11:30:11
1142595617242	2006-03-17 03:40:17 -0800	382	20.7883, 0.00005, 0.024, 01 Jan 1996, 11:40:10
1142596218161	2006-03-17 03:50:18 -0800	383	20.7513, 0.00005, 0.022, 01 Jan 1996, 11:50:10
1142596817278	2006-03-17 04:00:17 -0800	384	20.7091, 0.00005, 0.022, 01 Jan 1996, 12:00:11
1142597417478	2006-03-17 04:10:17 -0800	385	20.6558, 0.00005, 0.022, 01 Jan 1996, 12:10:10
1142598017207	2006-03-17 04:20:17 -0800	386	20.6273, 0.00005, 0.023, 01 Jan 1996, 12:20:10
1142598617243	2006-03-17 04:30:17 -0800	387	20.5311, 0.00005, 0.024, 01 Jan 1996, 12:30:11
1142599217342	2006-03-17 04:40:17 -0800	388	20.5244, 0.00005, 0.025, 01 Jan 1996, 12:40:11
1142599817432	2006-03-17 04:50:17 -0800	389	20.4530, 0.00005, 0.023, 01 Jan 1996, 12:50:11
1142600417333	2006-03-17 05:00:17 -0800	390	20.3890, 0.00005, 0.024, 01 Jan 1996, 13:00:10
1142601017301	2006-03-17 05:10:17 -0800	391	20.3397, 0.00005, 0.025, 01 Jan 1996, 13:10:09
1142601617418	2006-03-17 05:20:17 -0800	392	20.3128, 0.00005, 0.031, 01 Jan 1996, 13:20:09
1142602217237	2006-03-17 05:30:17 -0800	393	20.2405, 0.00005, 0.033, 01 Jan 1996, 13:30:10
1142602817253	2006-03-17 05:40:17 -0800	394	20.7132, 0.00005, 0.034, 01 Jan 1996, 13:40:10
1142603420751	2006-03-17 05:50:20 -0800	395	20.9352, 0.00005, 0.035, 01 Jan 1996, 13:50:10
1142604017527	2006-03-17 06:00:17 -0800	396	21.1775, 0.00005, 0.035, 01 Jan 1996, 14:00:10
1142604617288	2006-03-17 06:10:17 -0800	397	21.5304, 0.00005, 0.035, 01 Jan 1996, 14:10:11
1142605217292	2006-03-17 06:20:17 -0800	398	21.3885, 0.00005, 0.036, 01 Jan 1996, 14:20:11
1142605817232	2006-03-17 06:30:17 -0800	399	21.3123, 0.00005, 0.037, 01 Jan 1996, 14:30:10
1142606417351	2006-03-17 06:40:17 -0800	400	21.6514, 0.00005, 0.038, 01 Jan 1996, 14:40:10
1142607017291	2006-03-17 06:50:17 -0800	401	21.5593, 0.00005, 0.038, 01 Jan 1996, 14:50:10
1142607617273	2006-03-17 07:00:17 -0800	402	21.4343, 0.00005, 0.038, 01 Jan 1996, 15:00:11
1142608217491	2006-03-17 07:10:17 -0800	403	21.4092, 0.00005, 0.038, 01 Jan 1996, 15:10:10
1142608817302	2006-03-17 07:20:17 -0800	404	21.3881, 0.00005, 0.038, 01 Jan 1996, 15:20:10
1142609417532	2006-03-17 07:30:17 -0800	405	21.6112, 0.00005, 0.039, 01 Jan 1996, 15:30:11
1142610017298	2006-03-17 07:40:17 -0800	406	21.4691, 0.00005, 0.039, 01 Jan 1996, 15:40:11
1142610617329	2006-03-17 07:50:17 -0800	407	21.4292, 0.00005, 0.040, 01 Jan 1996, 15:50:11
1142611217208	2006-03-17 08:00:17 -0800	408	21.4044, 0.00005, 0.040, 01 Jan 1996, 16:00:10
1142611817232	2006-03-17 08:10:17 -0800	409	21.7756, 0.00005, 0.041, 01 Jan 1996, 16:10:09
1142612420621	2006-03-17 08:20:20 -0800	410	21.5259, 0.00005, 0.041, 01 Jan 1996, 16:20:09
1142613017248	2006-03-17 08:30:17 -0800	411	21.4641, 0.00005, 0.041, 01 Jan 1996, 16:30:10
1142613617332	2006-03-17 08:40:17 -0800	412	21.4108, 0.00005, 0.042, 01 Jan 1996, 16:40:10




Chart was not created because line would be flat with value of 5.0E-5

