

2008-06-03 14:06:26 LPA2-W #005114

TEST NUMBER 1

TEST REF C903 RV

TEST TYPE: Normal Test

NAS CLASS:-

6

μm	5-15	15-25	25-50	50-100	100+
NAS	6	3	3	00	00

ISO CODE 16/14/9

SAMPLE VOLUME 15mL

μm	/100ml
5-15	8555
15-25	287
25-50	44
50-100	0
100+	0

2008-06-03 19:17:02 LPA2-W #005114

TEST NUMBER 2

TEST REF C903 RV

TEST TYPE: Normal Test

NAS CLASS:-

7

μm	5-15	15-25	25-50	50-100	100+
NAS	6	6	7	00	00

ISO CODE 16/14/12

SAMPLE VOLUME 15mL

μm	/100ml
5-15	8508
15-25	2697
25-50	552
50-100	0
100+	0

2008-06-03 19:22:37 LPA2-W #005114

TEST NUMBER 3

TEST REF C903 RV

TEST TYPE: Normal Test

NAS CLASS:-

5

μm	5-15	15-25	25-50	50-100	100+
NAS	5	4	5	00	00

ISO CODE 15/13/10

SAMPLE VOLUME 15mL

μm	/100ml
5-15	4173
15-25	686
25-50	199
50-100	0
100+	0

2008-06-04 14:15:09 LPA2-W #005114

TEST NUMBER 4

TEST REF C903 RV

TEST TYPE: Normal Test

NAS CLASS:-

6

μm	5-15	15-25	25-50	50-100	100+
NAS	6	4	5	00	00

ISO CODE 15/14/10

SAMPLE VOLUME 15mL

μm	/100ml
5-15	8838
15-25	598
25-50	132
50-100	0
100+	0