

CETACEAN RESEARCH TECHNOLOGY

4728 12TH AVENUE NE SEATTLE, WASHINGTON 98105-4402 (206) 297-1310
CRTINFO@CETRESTEC.COM WWW.CETRESTEC.COM




C55 HYDROPHONE WIRING INSTRUCTIONS

Power your C55 hydrophone with any voltage ranging from 5V to 32V DC.

- Pin 1) **White wire** ➤ hydrophone output signal. The output signal will have a DC voltage offset equal to approximately one half of the power supply voltage. You should place a blocking capacitor between this wire and the input of your recording device. The low frequency limit of the hydrophone will be dictated by the input impedance, R , of your recording device and the capacitance, C , of the blocking capacitor. Low frequency signals will begin to roll off at a frequency $f = 1/(2\pi RC)$.
- Pin 2) **Red wire** ➤ hydrophone power. Apply between 5V and 32V DC to this wire. ***Power must come from a battery or from a regulated supply with no more than 0.5V ripple.***
- Pin 3) **Black wire & shield drain wire** ➤ common. This is power, signal, and shield common, and is connected to the stainless steel hydrophone connector shell. Connect black wire to the ground on your data acquisition or recording system. Leave shield drain wire disconnected, unless connecting it to ground improves noise rejection.

Wiring connections must be made at static-safe workstation.

Pin Configuration:



Male Face View

Your hydrophone has the following specifications.

Model Number: **C55**
Serial Number: **678**
Purchase Date: 05 April 2013
Transducer Sensitivity: -184.7dB, re. 1V/ μ Pa
Preamplifier Gain: 20.30dB
Effective Sensitivity: -164.4dB, re. 1V/ μ Pa
Linear Bandwidth: 20Hz to 44kHz (± 3 dB)
Useable Bandwidth: 9Hz to 100kHz (+3/-12dB)
Power Supply: 5 to 32 Vdc
Overload Acoustic Pressure: 169 to 186dB, re 1 μ Pa
Mating Connector: Subconn MCIL3F

Cetacean Research™ hydrophones carry a 90-day limited warranty. Defective units will be repaired or replaced at the manufacturer's discretion. Assembly at other than a static-safe workstation, mishandling, or damage to the cable or connector will void the warranty.