


QA INSPECTION REPORT

REPORT NO:

MODEL	901209-1 REV.2	ORDER NO.			INSP.DATE	2016-4-18	
CUSTOMER		LOT SIZE			SAMPLE SIZE		
INSP.ITEM		SPECIFICATION	INSP.RECORD				
SURFACE QUALITY		40-20	1 40-20	2 40-20	3	4	5
			6	7	8	9	10
EDGE CHIPS			1	2	3	4	5
			6	7	8	9	10
PROTECTIVE CHAMFER		0.5*45° MAX	1 0.5*45°	2 0.5*45°	3	4	5
			6	7	8	9	10
RINGS/IRR		R1:8/2 R2:8/1	1 8/2 8/1	2 8/2 8/1	3	4	5
			6	7	8	9	10
OUTSIDE DIAMETER		250.00+0.00/-0.10	1 249.96	2 249.98	3	4	5
			6	7	8	9	10
CENTER THICKNESS		22.00±0.20	1 22.09	2 22.12	3	4	5
			6	7	8	9	10
CENTERING		2'	1 2'	2 2'	3	4	5
			6	7	8	9	10
CLEAR APERTURE			1	2	3	4	5
			6	7	8	9	10
EFL			1	2	3	4	5
			6	7	8	9	10
OTHERS	1		1	2	3	4	5
			6	7	8	9	10
	2		1	2	3	4	5
			6	7	8	9	10
	3		1	2	3	4	5
			6	7	8	9	10
	4		1	2	3	4	5
			6	7	8	9	10
TRANSMISSION							
REMARK							
INSP.RESULT		☒ PASSED ☐ REJECTED					
INSPECTOR		Fang Wei		APPROVAL Xu Yu			

Date: 2015-11-21

Time: 10:58:03 下?



Description: ReflCorr : C:\UVWINLAB\DATA\901209-1-REV2.sp

Spectrum Name: REFLC0.SP

Scan Speed: 480.00 nm/min

Date Created: Sat Nov 21 10:58:44 2015

Slit Width: 1.0000 nm

Instrument Model: Arithmetic

Smooth Bandwidth: 2.00 nm

Data Interval: 1.0000 nm