

PO Box 518
620 Applegate St.
Philomath OR 97370
wetlabs@wetlabs.com



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

Date: 7/14/17

Customer: MBARI

S/N# BB2FLMBC

Technician: KM

Repairs and Modifications:

No problems found with incoming instrument. Completed standard Service & testing. Calibrated the instrument. New CHL Charsheet, 470nm Charsheet, 650nm Charsheet, device file and other information included on cd.