



Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone
Fax

+1-425-643-9866
+1-425-643-9954
www.seabird.com

SERVICE REPORT

Service Request
Date
Sales Order

1005504438
24-APR-2018
315047440

CUSTOMER INFORMATION

Name: MBARI - MONTEREY BAY AQUARIUM RESEARCH INSTITUTE
Account : 40279834
CHRIS WAHL
CWAHL@MBARI.ORG
831 775 1746

PO Number:

Bill To Address

ACCOUNTS PAYABLE;7700 SANDHOLDT RD.;
MOSS LANDING,CA,95039-9644,US

Ship To Address

7700 SANDHOLDT ROAD;BUILDING A;
MOSS LANDING,CA,95039-9644,US

PRODUCT INFORMATION

Item: GPCTD.LEGACY

Item Description: (LEGACY) Glider Payload CTD

Serial: 0118

Special Notes

Services Requested:
Evaluate/Repair Instrumentation.
Perform Routine Calibration Service.
Replace the instruments "O"-rings.
Perform hydrostatic pressure test.

Problems Found:
No problems found

Services Performed:
Performed initial diagnostic evaluation.
Performed "POST" cruise calibration.
Performed pressure calibration.
Replaced the O-rings.
Performed a hydrostatic pressure test.
Performed complete system check and full diagnostic evaluation.
Installed NEW AF24173 Anti-foulant cylinder(s).

Item	Item Description	Qty
REPLACEAF	Extra charge to install one antifoulant device, includes one 801542.1.	1
CAL_PAYLOAD	CALIBRATE GPCTD CONDUCTIVITY AND TEMPERATURE SENSORS (FRRF)	1
PCAL_PAYLOAD	CALIBRATE GPCTD PRESSURE SENSOR (FRRF)	1
CNCRTGPCTD	CONFIRM AND RE-CERTIFY PAYLOAD CTD (FRRF)	1

Unbilled Items

Item	Item Description	Qty
------	------------------	-----



Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone
Fax

+1-425-643-9866
+1-425-643-9954
www.seabird.com

SERVICE REPORT

Service Request

Date

Sales Order

1005504438

24-APR-2018

315047440

801542.1	AF24173 ANTI-FOULANT, SINGLE CYLINDER, V2	1
----------	---	---



Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone
Fax

+1-425-643-9866
+1-425-643-9954
www.seabird.com

SERVICE REPORT

Service Request

Date

Sales Order

1005504438

24-APR-2018

315047440

PRODUCT INFORMATION

Item: 43F.LEGACY

Item Description: (LEGACY) SBE 43F Oxygen Sensor

Serial: 43-0212

Special Notes

Services Requested:

Evaluate/Repair Instrumentation.

Perform Routine Calibration Service.

Problems Found:

No problems found

Services Performed:

Performed initial diagnostic evaluation.

Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_43	SBE 43 dissolved oxygen sensor calibration	1

Unbilled Items

Item	Item Description	Qty



SERVICE REPORT

Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone
Fax

+1-425-643-9866
+1-425-643-9954
www.seabird.com

Service Request

Date

Sales Order

1005504438

24-APR-2018

315047440

PRODUCT INFORMATION

Item: 05M.LEGACY

Item Description: (LEGACY) SBE 05M Submersible Pump

Serial: 051537

Special Notes

Services Requested:

Evaluate/Repair Instrumentation.

Problems Found:

No problems found

Services Performed:

Performed initial diagnostic evaluation.