



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA

Website: <http://www.seabird.com>

Phone: (425) 643-9866

FAX: (425) 643-9954

Email: seabird@seabird.com

SBE Pressure Test Certificate

Test Date: 11/4/2013 Description SBE-43F DO Sensor

SBE Sensor Information:

Model Number: 43f

Serial Number: 0212

Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For 15 Minutes

High Pressure Test: 870 PSI Held For 30 Minutes

Passed Test: ☒

Tested By: RH

