



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA
www.seabird.com

Phone: (425) 643-9866

Fax: (425) 643-9954

Email: seabird@seabird.com

Pressure Test Certificate

Test Date: **2/2/2016**

Description: **Payload Glider CTD**

Sensor Information:

Model Number: **Payload** **40**

Serial Number: **0173** **500**

Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** Minutes

High Pressure Test: **40** PSI Held For: **15** Minutes

Passed Test: **Yes**

Tested By: **VG**

High pressure is generally equal to the maximum depth rating of the instrument

