

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0206
CALIBRATION DATE: 15-Feb-17

Glider Payload CTD CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.911667e-001
h = 1.263606e-001
i = -6.150068e-005
j = 2.158357e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 3.5703e-007

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2800.73	0.00000	0.00000
1.0000	34.7312	2.96938	5591.20	2.96939	0.00001
4.4999	34.7120	3.27584	5802.61	3.27584	-0.00000
15.0000	34.6705	4.25562	6431.24	4.25560	-0.00001
18.5000	34.6617	4.60007	6637.84	4.60008	0.00001
24.0000	34.6524	5.15693	6958.56	5.15695	0.00001
29.0000	34.6479	5.67783	7245.38	5.67783	-0.00001
32.5000	34.6456	6.04958	7443.12	6.04954	-0.00005

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

