

MBARI Carbon WG Console Help

Command syntax help is organized using the following topics:

- REG built-in register variables
 - QUEUE sample queue
 - FSYS file system
 - DIAG diagnostics/debug
 - GET get configuration/state/IO
 - SET set configuration/state/IO
 - CLR clear IO
 - PULSE momentary IO cycling
 - LI licor control
 - SCR script control
 - MAC macros
 - REF reference
-

Group [REG]

CREG

```
use : creg [<reg>] [op <val>]
brief : show, manipulate counter reg (CX0:SCR_CX_COUNT)
input : reg - register [CXn|*]
input : op - operator [=,--,++,+.=.-=.*=,/=,%=]
input : val - value [int32]
return : requested value(s)
```

DREG

```
use      : dreg [<reg>] [op <val>]
brief    : show, manipulate double reg (DX0:SCR_DX_COUNT)
input    : reg   - register [DXn|*]
input    : op    - operator [=,--,++,+.=.-=.*=,/=]
input    : val   - value [double]
return   : requested value(s)
```

FREG

```
use      : freg [<reg>] [op [<val>]]
brief    : show, manipulate flag(s) (0:31)
input    : reg   - register [FXn|*]
input    : op    - operator [=,!]
input    : val   - value [1|0]
return   : requested value(s)
```

Group [QUEUE]

QCOUNT

```
use      : qcount
brief    : return queue entries
return   : number of queue entries
```

QCLEAR

```
use      : qclear
brief    : delete all queue entries
return   : none
```

QPEEK

```
use      : qpeek [<first>] [<n>]
brief    : examine queue entries (don't delete)
input    : first - starting queue entry
input    : n     - number of entries or '*'
return   : decoded queue entries
```

QPOP

```
use      : qpop [Q] <n>
brief    : delete and optionally show next (oldest) n entries from queue
input    : n - number of entries or '*' for all
input    : Q - suppress output (just delete)
return   : oldest n entries in the queue (or none if Q specified)
```

QSHOW

```
use      : qshow
brief    : print queue summary
return   : queue summary
```

Group [FSYS]

CAT

```
use      : cat <file>
brief    : show file contents
input    : file - file to show
return   : file contents
```

CFGEX

```
use      : cfgex <file>
brief    : export current configuration to file
input    : file - destination file
          [NOTE : file will be overwritten if it exists]
return   : none
```

CP|MV

```
use      : cp <src> <dest>
brief    : copy contents from src to dest
input    : src  - file to copy FROM
input    : dest - file to copy TO
return   : none
```

LC

```
use      : lc <file>
brief    : count lines in file
          [any console input interrupts]
input    : file - file to evaluate
return   : number of lines, execution time
```

LESS

```
use      : less <file> [<lines>]
brief    : page a text file to the console (with search)
input    : file  - file to evaluate
input    : lines - number of text lines per page
return   : none
notes    :
```

Command Summary

```
Move      : Up [u U]   Down [d D CR]
Jump      : First [g <] Last [G >]
Search    : Find [f /] Next [n N]
Session   : Help [h H ?] Quit [q Q x X]
```

LOG

```
use      : log <msg>
brief    : add log annotation
input    : msg - message to log
return   : none
```

LS|DIR

```
use      : ls [<dir>]
brief    : list file info
input    : dir - directory [default: root, i.e. 0:/]
return   : list of files in directory
```

LSEG

```
use      : lseg <hnd> [<val>]
brief    : get/set log segment size
input    : hnd - log handle 0:syslog 1:data 2:raw
input    : val - log segment size, n>0
           may also use K,M,G suffix (e.g. 10M) for KByte, MByte, GByte
return   : none
```

LSHOW

```
use      : lshow [<hnd>]
brief    : show log summary
input    : hnd - log handle 0:syslog 1:data 2:raw
return   : none
```

REN

```
use      : ren <from> <to>
brief    : change file name
input    : from - current file name
input    : to   - new file name
return   : none
```

RM

```
use      : rm <file> [<file>...]
brief    : delete one or more files
input    : file - file to delete
return   : none
```

STATS

```
use      : stats <vol>
brief    : show file system information
input    : vol - volume
return   : file system summary
```

UPLOAD

```
use      : upload <dest>
brief    : upload ASCII file to root directory
input    : dest - file name
           dest will be overwritten if it exists
           otherwise, it will be created
           upload times out after 4 s, or may be terminated using CTRL_D
           [NOTE : not a production feature;
           some terminal emulators may not support direct ASCII upload]
return   : none
```

Group [DIAG]

HELP|?

```
use      : help [v]
brief    : show help info
input    : v - verbose output (this file) (optional)
return   : help information
```

HIST|!

```
use      : hist [v] [<id>]
brief    : show command history, execute previous command
input    : v      - verbose output (optional, ignored if id specified)
input    : <id> - command ID (integer) (optional)
return   : command history by default, executes queued command if specified
note     : up/down arrow keys may be used to traverse the queue
```

CON

```
use      : con <port>
brief    : open console to device
          [terminate session using CTRL_D]
input    : port - serial port [LI|HO]
return   : none
```

DEBUG

```
use      : debug [<op><mask>]
brief    : show/set per-module debug output flags
          MOD_ACT =0x0001 // cli_imp1
          MOD_TMR =0x0002 // timers
          MOD_SCR =0x0004 // script
          MOD_LIC =0x0008 // licor
          MOD_SER =0x0010 // ioserial
          MOD_CFG =0x0020 // config
          MOD_MAIN=0x1000 // main
          MOD_TOKS=0x2000 // command token

input    : op - operator [+ :set - :clear = :assign]
input    : mask - hex bit field [0-FFFFFFFF]
return   : debug flags
examples :
  >>debug =0
  SYS_MDFLAGS=x00000000
  >>debug +4
  SYS_MDFLAGS=x00000004
  >>debug +8
  SYS_MDFLAGS=x0000000c
  >>debug -8
  SYS_MDFLAGS=x00000004
```

PMEM

use : pmem brief : show memory pool contents return : memory pool contents

PRINT, PRINTLN

```
use      : print <fmt> [<var>...]
brief    : print formatted messages to console using variable expansion escape
          character support.
          fmt is a quoted format string, which may include , escaped variables
          (with $ prefix), printf formats
```

```
var are variables (without $ prefix)
println variant automatically appends CRLF to string
return : formatted console message (upper case)
examples :
    >>CREG CX0 = 20
    // print CX using default (hex) format
    >>PRINT "$CX0\r\n"
    X00000014
    // print CX using alternative (decimal) format
    >>PRINT "%ld\r\n" CX0
    20
```

ECHO

```
enable using WITH_UI_ECHO
use : echo [<op><dest>] [<msg>...]
brief : print messages using variable expansion escape character support
input : op - operator [+enable -:disable]
input : dest - destination [C:console H:host]
input : msg - text, variables, registers
        quoted strings preserve space, unquoted strings are space-delimited
        variables and registers are denoted by $ (e.g. $verbose, $CX2)
        escape characters are supported in quoted strings [\r\n\t\v\xDD]
return : expanded message
```

VER

```
use : ver
brief : show firmware version
return : version information
```

UPTIME

```
use : uptime
brief : return elapsed time since last power cycle
return : time since last power cycle
```

RESET

```
use      : reboot
brief    : reboot the firmware
return   : none
```

HELLO

```
use      : hello [<msg>...]
brief    : print test messages to console
          [deprcated; use PRINT, PRINTLN is recommended]
input    : msg - text, variables, registers
          quoted strings preserve space, unquoted strings are space-delimited
          variables and registers are denoted by $ (e.g. $verbose, $CX2)
          escape characters are supported in quoted strings [\r\n\t\v\xDD]
return   : 'hello' [args...]
```

SWEEP

```
use      : sweep <delay_ms>
brief    : turn all IO bits on and then off in order
input    : delay_ms - switching delay
return   : sets all IO bits LO (except serial transceivers)
```

PROTO

```
use      : proto (args)
brief    : runs selected prototype code [experts only]
          enabled code depends on compile-time configuration in cwgsys.h
input    : args - arguments (varies)
return   : varies
```

Group [GET]

GET AD

```
use      : get ad <port> [DXn]
brief    : show A/D converter voltage
input    : port - ADC port [0:1]
output   : DXn - store result in DXn [0 < n < SCR_DX_COUNT-1] (optional)
return   : ADC - voltage [0-5V]
```

GET BITS

```
use      : get bits [CXn]
brief    : show IO bit state summary
output   : CXn - store result in CXn [0 < n < SCR_CX_COUNT-1] (optional)
return   : IO bit states
```

GET NEXT

```
use      : get next
brief    : show time (s) until next scheduled activity
return   : time (s) or SYS_MAX_SLEEP_MS if no events pending
```

GET SER

```
use      : get ser <port>
brief    : read from serial port string until buffer full or end of input
input    : port - serial port [port 0:2]
return   : bytes read
```

GET TIME

```
use      : get time [u]
brief    : show system time
input    : u - output epoch time (optional)
return   : current date/time
```

GET VAR get var []

```
use      : get var [V] [key...]
brief    : display one or more variable values
input    : V    - verbose output
input    : key - variable name (key)
return   : value of <key>, or list of key/value pair
          values if if unspecified
```

Group [SET]

SET BAUD

```
use      : set baud <port> <bps>
brief    : set UART communication speed (bps)
input    : port - serial port [0:2|H|C|L]
input    : bps  - rate (bps) [300,600,1200,2400,4800,9600,19200,38400]
return   : none
```

SET BITS

```
use      : set bits [id...]
brief    : set (HI) one or more digital IO bits
input    : id - bit number or name
          #n : bit n (decimal)
              0:19 valves/acid pump/sea pump
              20:21 pump
              22:22 licor
              23:24 analog
              26:28 RS232 drivers
          xn : bit mask n [0-FFFFF] (hex)
          an : ADC n 0:1
          a* : all ADC
          vn : valve n A&B
          vnA|B : valve n A,B
          v* : all valves
          pn : DC pump n 0:3
          pi,pr : air pump
          pe,pw : seawater pump
          pa : acid pump
          ps : sample pump
          pd : DC pumps (air, seawater)
```

```
px : solenoid pumps (acid, sample)
p* : all pumps
  l  : licor DC power
  uh : host xcvr
  ul : licor xcvr
  uc : console xcvr
return : none
```

SET SER

```
use      : set ser <port> <string>
brief    : send serial string to specified serial port
input    : port - serial port [0:2|H|C|L]
input    : string - string to send
return   : none
```

SET TIME get var []

```
use      : set time YY[YY] MM DD hh mm ss
brief    : set system time/date
input    : YY[YY] - year    [00-9999]
input    : MM      - month  [01:12]
input    : DD      - day    [01:31]
input    : hh      - hour   [00:24]
input    : mm      - minute [00:59]
input    : ss      - second [00:59]
return   : new time
```

SET VAR

```
use      : set var <key> <val>
brief    : set variable <key> to <val>
input    : key - variable name
input    : val - new value (must be appropriate type and format)
return   : none
```

Group [CLR]

CLR BITS

```
use      : clr bits [id...]
brief    : clear (LO) one or more digital IO bits
input    : id      - bit number or name (see SET BITS)
return   : none
```

Group [PULSE]

PULSE BITS

```
use      : pulse bits <dir> <delay_ms> bits [id...]
brief    : pulse one or more digital IO bits
input    : dir      - direction [L|0|H|1]
input    : delay_ms - pulse duration (ms)
input    : id       - bit number or name (see SET BITS)
return   : none
```

PULSE NBITS

```
use      : pulse nbits <dir> <ta> <tb> <n> bits [id...]
brief    : pulse one or more digital IO bits w/ specified duty cycle, number of
cycles
input    : dir      - direction [L|0|H|1]
input    : ta       - pulse duration (dir, ms)
input    : tb       - pulse duration (!dir, ms)
input    : cycles   - pulse duration (ms)
input    : id       - bit number or name (see SET BITS)
```

Group [LI] (licor)

LI CAL

```
use      : li cal <id> [<spanc> D<delay>]
brief    : perform calibration
input    : id      - cal ID [z:zero s:span s2:span2]
input    : spanc   - span concentration (required for span cal)
input    : delay   - extend delay (in addition to 10s max implemented)
return   : none
```

LI INIT

```
use      : li init
brief    : initialize licor
return   : none
```

LI NMEAS

```
use      : li nmeas <handler> <sample_n> <period> [options]
brief    : take multiple measurements, process using specified handler
           send output to specified destination(s).
input    : handler  - handler [a:average b:baseline r:]
input    : sample_n - number of records to process
input    : period   - licor output period (s) [>=0.5]
input    : options  - options
                   O<destination> : H:host C:console L:log A:all
                   R<rec_type>    : character
                   C<rec_context> : character
                   T<tmout>       : timeout (ms) default 1.1*sample_n*period
return   : record (varies by handler)
```

LI OR

```
use      : li or <period>
brief    : set licor output rate
input    : period - output period (s) [period >=0.5]
return   : none
```

LI STAT

```
use      : li stat [i<indent> v] <tag>...]
brief    : output licor status to console (formatted xml)
input    : indent - number of space to indent
input    : v      - verbose output
input    : tag    - output contents of first occurrence of tag
return   : none
examples :
  // show <rs232>...</rs232> and <cfg>...</cfg>
  li stat rs232 cfg
  // show all, indent 8 spaces
  li stat i8
```

LI TM

```
use      : li tm [options]
brief    : take one or more measurements using specified delay.
          Send output to specified destination(s)
input    : options
          O<destination> : H:host C:console L:log A:all (default:C)
          R<rec_type>    : character (default:U)
          C<rec_context> : character (default:?)
          N<samples>     : number of samples (default:1)
          D<delay>       : delay between samples (ms) (default:0)
return   : record
```

Group [SCR] (script)

SCR SDEBUG

```
use      : scr sdebug +|- <id>
brief    : enable/disable debug output for specified script
input    : id - script name or handle
return   : none
```

SCR DELAY

```
use      : scr delay <time_ms> <id>
brief    : delay specified period before executing next command
input    : time_ms - delay time (ms)
input    : id      - script name or handle
return   : none
```

SCR END

```
use      : end
brief    : end of script [only valid in script context]
return   : none
```

SCR JOG

```
use      : scr jog <id>
brief    : single-step the specified script [enters manual mode]
input    : id - script name or handle
return   : none
```

SCR LOAD

```
use      : scr load [o<+/->] <file> [<file>...[o<+/->]...]
brief    : load specified script(s) into next available handle.
input    : file - script file name
input    : o    - enable/disable overwrite for files that follow (may use more than
once)
return   : confirmation or error message
```

SCR RUN

```
use      : scr run <id>
brief    : run the specified script.
input    : id - script name or handle
return   : none
```

SCR RESET | KILL

```
use      : scr reset <id>
brief    : reset the specified script. Does not change hardware states.
input    : id - script name or handle
return   : none
```

SCR SCH

```
use      : scr sch <period_ms> [<sync>] <id>
brief    : schedule script, optionally synchronize (run)
input    : period_ms - schedule period (ms)
input    : sync      - sync now [Y|N]
input    : id        - script name or handle
return   : none
```

SCR SHOW

```
use      : scr show [<id>...]
brief    : print script parameter summary for one or more scripts
input    : id - script name or handle or '*'
return   : none
```

SCR SYNC

```
use      : scr sync <time_ms>
brief    : start script after optional delay
           [shifts scheduled start time]
input    : time_ms - delay time (ms) (optional)
return   : none
```

SCR UNLOAD

```
use      : scr unload <id>
brief    : unload specified script.
input    : id - script name or handle
return   : none
```

Group [MAC] (macro)

Use notes

MDIO and MMEAS separate IO manipulation from measurement processing. MDIO may be used to set up valves and pumps with specified delays. MMEAS is used to process a set of licor measurements, using specified parameters. They provide fine-grained control of cycle operations, as well as simple presets (macros) that encapsulate several IO operations and timing, with minimal parameter overrides.

The use of M macros are deprecated; use MDIO, MMEAS.

M equivalents are implemented using mdio and mmeas:

```
PURGE:
    mdio purge
FILL:
    mdio fill
ACID:
    mdio acid
ZERO:
    mdio zero
MEAS:
    mdio sample
    mmeas prebl sample postbl
    mdio measx
AIR:
    mdio air
    mmeas air
    mdio measx
STD:
    mdio std
    mmeas std
    mdio measx
SZERO:
    mdio szero
REST:
    mdio rest
```

MDIO

```
use      : mdio [ vd<dir> vt<ton> ph<ton> pl<toff> pn<cycles>
               <vset> <pump> +sea -sea +air -air +li -li +v -v <preset> d<delay>
show]
brief    : Assert digital IO and macros w/ optional delay between operations.
           Operations are performed left to right
```

General:

input : d - delay time (msec)
input : +/-v - verbose output on/off
input : show - show current settings, defaults

Manual Pump/Valve Control [1]:

input : vd - valve dir:H|L
input : vt - valve pulse time (msec)
input : ph,l - pump pulse time (msec)
input : pn - pump cycles

Macro Control [1]:

input : mt - macro (preset) valve pulse time (msec)
input : mh,l - macro (preset) pump pulse time (msec)
input : mn - macro (preset) pump cycles

Operation Presets [2,3]:

input : +/-sea - seawater pump on/off
input : +/-air - air pump on/off
input : +/-li - licor on/off

input : <vset> - switch valve set :

iofill ioair iopurge iozero ioszero iostd iostdfill iomeas iorest

input : <preset> - macro presets :

fill air purge zero szero std stdfill sample measx rest

input : <pump> - cycle pump :

psample pacid

[1] vt, ph, pl, pn, mt, mh, ml, mn

- are sticky (per command line), may set multiple times
- use a default value if unset
- defaults used if unset (see cli_impl.h)
- vt, ph, pl, pn do not apply to macro presets
- mt, mh, ml, mn apply only to macro presets

[2] measx turns off AIR_PUMP and LICOR after sampling

[3] pn and pump bits (pacid and/or psample) must be set to activate pump

pump bits accumulate (are OR'd) until pn specified (e.g. mdio pacid psample pn3)

pn may be set before or after pacid/psample
pn, mn are sticky (per command line), may set multiple times
pump bits (pacid, psample) reset on execution

MMEAS

```
use      : mmeas [r<rtype> c<ctx> p<period> h<handler> o<output>
              t<mtime> n<mcount> w<cal_wr> d<delay> meas <preset> ]
brief    : licor measurement cycle
input    : rtype      - record type (char) [UVBRS?]
input    : ctx        - measurement context (char) [UZSMasz?]]
input    : mtime      - measurement time (msec)
input    : mcount     - measurement count (alternative to mtime)
input    : handler    - measurement handler (A:average|B:baseline)
input    : period     - licor sample period (sec >=0.5)
input    : output     - output (A:all H:host C:console L:log)
input    : cal_wr     - enable/disable writing calibration to licor [+|-]
input    : delay      - delay (msec)
input    : meas       - take measurement using current settings
input    : preset     - packaged settings: zero air std sample prebl postbb1 zcal
scal
```

Group [REF] (reference)

IO bit indices

Name	bit	hex	binary
SV0A	00	0x00000001	0000 0000 0000 0000 0000 0000 0000 0001
SV0B	01	0x00000002	0000 0000 0000 0000 0000 0000 0000 0010
SV1A	02	0x00000004	0000 0000 0000 0000 0000 0000 0000 0100
SV1B	03	0x00000008	0000 0000 0000 0000 0000 0000 0000 1000
SV2A	04	0x00000010	0000 0000 0000 0000 0000 0000 0001 0000
SV2B	05	0x00000020	0000 0000 0000 0000 0000 0000 0010 0000
SV3A	06	0x00000040	0000 0000 0000 0000 0000 0000 0100 0000
SV3B	07	0x00000080	0000 0000 0000 0000 0000 0000 1000 0000
SV4A	08	0x00000100	0000 0000 0000 0000 0000 0001 0000 0000
SV4B	09	0x00000200	0000 0000 0000 0000 0000 0010 0000 0000
SV5A	10	0x00000400	0000 0000 0000 0000 0000 0100 0000 0000

