

Certificate/Report of Calibration



241 Springside Drive
Akron, OH 44333
Phone: 330-668-1444
www.dataq.com

Serial Number:	678674EA	Test Result:	PASS
Model:	DI-2108	Cal Date:	1/14/2025 14:39 PM
Hardware Revision:	R	Cal Due Date:	1/14/2026
Firmware Revision:	1.39	Manufacture Date:	1/14/2025
Procedure Name:	DATAQ Automated Calibration	Temperature:	23.64 Deg C
Procedure Revision:	1.13	Calibrated by:	G Zahradnick

DATAQ Instruments certifies that the above listed instrument meets or exceeds all specifications as stated in the referenced procedure (unless otherwise noted). It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology (NIST), or to NIST accepted intrinsic standards of measurement, or derived by the ration type of self calibration techniques.

Standards Used

Asset	Manufacturer	Model	Description	Cal Date	Due Date
8737202	FLUKE	FLUKE 525A	CALIBRATOR	8/1/2024	8/1/2025

Test Data

CHANNEL NUMBER	INPUT RANGE	INPUT VOLTAGE	PREVIOUS VALUE	CALIBRATED RESULT	ACCEPTED LOW LIMITS	ACCEPTED HIGH LIMITS	TEST STATUS
1	10.000	9.500 0.000	- -	9.4995 0.0008	9.4950 -0.0050	9.5050 0.0050	Pass Pass
2	10.000	9.500 0.000	- -	9.4995 0.0007	9.4950 -0.0050	9.5050 0.0050	Pass Pass
3	10.000	9.500 0.000	- -	9.5002 0.0006	9.4950 -0.0050	9.5050 0.0050	Pass Pass
4	10.000	9.500 0.000	- -	9.4998 0.0006	9.4950 -0.0050	9.5050 0.0050	Pass Pass
5	10.000	9.500 0.000	- -	9.4995 0.0003	9.4950 -0.0050	9.5050 0.0050	Pass Pass
6	10.000	9.500 0.000	- -	9.4999 0.0005	9.4950 -0.0050	9.5050 0.0050	Pass Pass
7	10.000	9.500 0.000	- -	9.4995 0.0006	9.4950 -0.0050	9.5050 0.0050	Pass Pass
8	10.000	9.500 0.000	- -	9.4995 0.0007	9.4950 -0.0050	9.5050 0.0050	Pass Pass

End of Test Data

Greg Zahradnick

For information about Re-Calibration, visit us online:
<https://www.dataq.com/support/calibration.html>