



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005 USA

Phone: +1 425-643-9866
Fax: +1 425-643-9954
E-mail: info@Sea-BirdScientific.com

SBE PH SENSOR CALIBRATION SHEET:

Date: February 11th, 2019

Sensor Serial Number: SBE C04

Calibration Date: Aug 23rd, 2018

f(P), 6th order polynomial:

f6	f5	f4	f3	f2	f1	f0
-7.3622E-22	5.7894E-18	-1.871E-14	3.2171E-11	-3.1889E-08	2.038E-05	-3.7606E-03

* f0 is not used in pH calculations.

k0 (HCl):

-1.4131

k2:

-1.1416E-03