

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

FLBB2K 2000-meter Certification

Date: January 11, 2017

S/N: FLBB2K-4592

Tester: Ross Greenley

Signature: Ross Greenley

This instrument has been pressure-tested to the depth specified below. WET Labs' certification of this instrument shall not apply to it if it has been repaired or altered by anyone other than WET Labs, or if it is subjected to misuse due to common negligence

WET Labs makes no warranty of any kind, express or implied, except that this instrument shall be of the standard quality of WET Labs, and buyer assumes all risk and liability resulting from the use of this instrument, whether used singly or in combination

WET Labs neither assumes nor authorizes any person to assume for WET Labs any other liability in connection with the sale or use of this instrument, and there are no oral agreements or warranties collateral to or affecting this disclaimer of warranty.

Test Protocol

Meter submerged to 2400 meters in three cycles; held at depth for a minimum of 12 hours. Certified to a depth of 2000 meters.