




THE INFORMATION CONTAINED IN THIS DRAWING IS THE SOLE PROPERTY OF MBARI. ANY REPRODUCTION IN PART OR WHOLE WITHOUT THE WRITTEN PERMISSION OF MBARI IS PROHIBITED.







Item No.	Qty	Part No.	Description	Material Specification
f	4		#0-80 x 1/8" panhead screws	Type 316 Stainless Steel
e	1		.087 x .020 O-ring	Viton, D75 (Apple Rubber)
d	1		.248 x .020 O-ring	Viton, D75 (Apple Rubber)
c	1		DuraFET Chip	Honeywell
b	1		7 X 1.5 mm O-ring	Silicone, D50
a	2		-011 O-ring	Viton, D75
7	1		DuraFET Cover	PEEK (natural)
6	1		DuraFET Mount	PEEK Rod (natural)
5	1		Silver Chloride Pellet	vanLondon 7.2 x 3.6 mm silver chlodide
4	1		Ref Pellet Housing	PEEK Rod (natural)
3	1		Isolation Washer	PEEK Rod (natural)
2	1		Counter Electrode	Type 316 Stainless Steel Rod
1b	1		Common Support (Config b)	PEEK Rod (natural)
1a	1		Common Support (Config a)	PEEK Rod (natural)
A2	1		DuraFET for PTChamber	
A1	1		Ref Cell for PTChamber	

Shop / Q.C.				UNLESS OTHERWISE SPECIFIED TOLERANCES ARE:		APPROVALS		DATE		Monterey Bay Aquarium Research Institute	
Inspector				1 PLACE DIM ±1/16	2 PLACE DIM ±.02	3 PLACE DIM ±.005	ANGLES ±1°	DRAWN	Hans Jannasch 5/30/11	MBARI 7700 Sandholdt RD. Moss Landing, CA 95039 PH: (831) 775-1700	
Shop Foreman				1. DIMENSIONS ARE IN INCHES				ENGINEER			
Engineer				2. DIMENSION LIMITS HELD AFTER PLATING.				CHECKED		Common DuraFET & Reference	
NEXT ASSY Description	NEXT ASSY Number			3. REMOVE BURRS AND SHARP EDGES .015 MAX.				PROJECT NO.	900918		
				4. MACHINED FILLET RADII .0150-.030.				PROJECT	Novel Sensors	SIZE B DWG. NO. 242jahaCommonDF+Ref REV. A	
				5. MACHINED DIAMETERS ON A COMMON CENTERLINE CONCENTRIC WITHIN .005				WEIGHT			
				6. PERPENDICULARITY & PARALLELISM OF MACHINED SURFACES .002 PER INCH. MAX .012 INCHES FOR SINGLE SURFACE.						SCALE 1:3 CAD FILE: Tornado/chem/SWfiles SHEET 1 OF 6	
Application				DO NOT SCALE DRAWING							

THE INFORMATION CONTAINED IN THIS DRAWING IS THE SOLE PROPERTY OF MBARI ANY REPRODUCTION IN PART OR WHOLE WITHOUT THE WRITTEN PERMISSION OF MBARI IS PROHIBITED.



REVISIONS			
REV.	DESCRIPTION	DATE	APPROVED
	INITIAL RELEASE		
A	Modified DFmount and CounterElectrode	9/15/11	hwj



1b	1		Common Support (Config b)	PEEK Rod (natural)	
1a	1		Common Support (Config a)	PEEK Rod (natural)	
Item No.	Qty	Part No.	Description	Material Specification	
Bill of Materials					
Shop / Q.C.		UNLESS OTHERWISE SPECIFIED TOLERANCES ARE:		APPROVALS DATE	
Inspector		1 PLACE DIM ±1/16	2 PLACE DIM ±.02	3 PLACE DIM ±.005	DRAWN Hans Jannasch 5/30/11
Shop Foreman		1. DIMENSIONS ARE IN INCHES			
Engineer		2. DIMENSION LIMITS HELD AFTER PLATING.			CHECKED
NEXT ASSY Description	NEXT ASSY Number	3. REMOVE BURRS AND SHARP EDGES .015 MAX.			PROJECT NO. 900918
		4. MACHINED FILLET RADII .0150-.030.			PROJECT Novel Sensors
		5. MACHINED DIAMETERS ON A COMMON CENTERLINE CONCENTRIC WITHIN .005			
		6. PERPENDICULARITY & PARALLELISM OF MACHINED SURFACES .002 PER INCH. MAX .012 INCHES FOR SINGLE SURFACE.			
Application	DO NOT SCALE DRAWING		WEIGHT		
Monterey Bay Aquarium Research Institute					
MBARI					
7700 Sandholdt RD.					
Moss Landing, CA 95039					
PH: (831) 775-1700					
Common DuraFET & Reference					
SIZE B		DWG. NO. 242jahaCommonDF+Ref		REV. A	
SCALE 1:3		CAD FILE: Tornado/chem/SWfiles		SHEET 2 OF 6	

THE INFORMATION CONTAINED IN THIS DRAWING IS THE SOLE PROPERTY OF MBARI ANY REPRODUCTION IN PART OR WHOLE WITHOUT THE WRITTEN PERMISSION OF MBARI IS PROHIBITED.







REVISIONS			
REV.	DESCRIPTION	DATE	APPROVED
	INITIAL RELEASE		
A	Modified DFmount and CounterElectrode	9/15/11	hwj



2		1		Counter Electrode		Type 316 Stainless Steel Rod	
Item No.		Qty		Part No.		Description	
						Material Specification	
Bill of Materials							
Shop / Q.C.		UNLESS OTHERWISE SPECIFIED TOLERANCES ARE:				APPROVALS DATE	
Inspector		1 PLACE DIM ±1/16	2 PLACE DIM ±.02	3 PLACE DIM ±.005	ANGLES ±1°	DRAWN Hans Jannasch 5/30/11	
Shop Foreman		1. DIMENSIONS ARE IN INCHES				ENGINEER	
Engineer		2. DIMENSION LIMITS HELD AFTER PLATING.				CHECKED	
		3. REMOVE BURRS AND SHARP EDGES .015 MAX.				PROJECT NO. 900918	
		4. MACHINED FILLET RADII .0150-.030.				PROJECT Novel Sensors	
NEXT ASSY Description	NEXT ASSY Number	5. MACHINED DIAMETERS ON A COMMON CENTERLINE CONCENTRIC WITHIN .005					
		6. PERPENDICULARITY & PARALLELISM OF MACHINED SURFACES .002 PER INCH. MAX .012 INCHES FOR SINGLE SURFACE.					
Application		DO NOT SCALE DRAWING				WEIGHT	
							
						MBARI 7700 Sandholdt RD. Moss Landing, CA 95039 PH: (831) 775-1700	
						Common DuraFET & Reference	
		SIZE B		DWG. NO. 242jahaCommonDF+Ref		REV. A	
		SCALE 1:3		CAD FILE: Tornado/chem/SWfiles		SHEET 3 OF 6	

THE INFORMATION CONTAINED IN THIS DRAWING IS THE SOLE PROPERTY OF MBARI ANY REPRODUCTION IN PART OR WHOLE WITHOUT THE WRITTEN PERMISSION OF MBARI IS PROHIBITED.

REVISIONS			
REV.	DESCRIPTION	DATE	APPROVED
	INITIAL RELEASE		
A	Modified DFmount and CounterElectrode	9/15/11	hwj



Note: Machine chamfer while turning pellet in clean soft collet with clean sharp tool. Use NO oil or other fluids. Handle pellet with clean latex gloves.

Note: Side with no foil is not machined (machined flat after epoxying into part #4)

SECTION F-F
SCALE 4 : 1

Item No.	Qty	Part No.	Description	Material Specification
5	1		Silver Chloride Pellet	vanLondon 7.2 x 3.6 mm silver chloride
4	1		Ref Pellet Housing	PEEK Rod (natural)
3	1		Isolation Washer	PEEK Rod (natural)

Bill of Materials											
Shop / Q.C.		UNLESS OTHERWISE SPECIFIED TOLERANCES ARE:				APPROVALS		DATE		Monterey Bay Aquarium Research Institute MBARI 7700 Sandholdt RD. Moss Landing, CA 95039 PH: (831) 775-1700 	
Inspector		1 PLACE DIM ±1/16	2 PLACE DIM ±.02	3 PLACE DIM ±.005	ANGLES ±1°	DRAWN Hans Jannasch		5/30/11			
Shop Foreman		1. DIMENSIONS ARE IN INCHES 2. DIMENSION LIMITS HELD AFTER PLATING. 3. REMOVE BURRS AND SHARP EDGES .015 MAX. 4. MACHINED FILLET RADII .0150-.030. 5. MACHINED DIAMETERS ON A COMMON CENTERLINE CONCENTRIC WITHIN .005 6. PERPENDICULARITY & PARALLELISM OF MACHINED SURFACES .002 PER INCH. MAX .012 INCHES FOR SINGLE SURFACE.				ENGINEER					
Engineer						CHECKED					
						PROJECT NO.		900918			
NEXT ASSY Description	NEXT ASSY Number					PROJECT		Novel Sensors			
Application		DO NOT SCALE DRAWING				WEIGHT					
						SIZE B		DWG. NO. 242jahaCommonDF+Ref		REV. A	
						SCALE 1:3		CAD FILE: Tornado/chem/SWfiles		SHEET 4 OF 6	

Monterey Bay Aquarium Research Institute
MBARI
7700 Sandholdt RD.
Moss Landing, CA 95039
PH: (831) 775-1700








Common DuraFET & Reference

SIZE B DWG. NO. 242jahaCommonDF+Ref REV. A

SCALE 1:3 CAD FILE: Tornado/chem/SWfiles SHEET 4 OF 6

THE INFORMATION CONTAINED IN THIS DRAWING IS THE SOLE PROPERTY OF MBARI ANY REPRODUCTION IN PART OR WHOLE WITHOUT THE WRITTEN PERMISSION OF MBARI IS PROHIBITED.

REVISIONS			
REV.	DESCRIPTION	DATE	APPROVED
	INITIAL RELEASE		
A	Modified DFmount and CounterElectrode	9/15/11	hwj



6		1		DuraFET Mount		PEEK Rod (natural)	
Item No.		Qty		Description		Material Specification	
Bill of Materials							
Shop / Q.C.		UNLESS OTHERWISE SPECIFIED TOLERANCES ARE:		APPROVALS		DATE	
Inspector		1 PLACE DIM ±.1/16	2 PLACE DIM ±.02	3 PLACE DIM ±.005	DRAWN Hans Jannasch		5/30/11
Shop Foreman		1. DIMENSIONS ARE IN INCHES		ENGINEER			
Engineer		2. DIMENSION LIMITS HELD AFTER PLATING.		CHECKED			
		3. REMOVE BURRS AND SHARP EDGES .015 MAX.		PROJECT NO.		900918	
		4. MACHINED FILLET RADII .0150-.030.		PROJECT		Novel Sensors	
NEXT ASSY Description	NEXT ASSY Number	5. MACHINED DIAMETERS ON A COMMON CENTERLINE CONCENTRIC WITHIN .005					
		6. PERPENDICULARITY & PARALLELISM OF MACHINED SURFACES .002 PER INCH. MAX .012 INCHES FOR SINGLE SURFACE.					
Application		DO NOT SCALE DRAWING		WEIGHT			


Monterey Bay Aquarium Research Institute

MBARI
7700 Sandholdt RD.
Moss Landing, CA 95039
PH: (831) 775-1700



Common DuraFET & Reference

SCALE	1:3	CAD	FILE: Tornado/chem/SWfiles	SHEET	5 OF	REV.	A
-------	-----	-----	----------------------------	-------	------	------	---



THE INFORMATION CONTAINED IN THIS DRAWING IS THE SOLE PROPERTY OF MBARI ANY REPRODUCTION IN PART OR WHOLE WITHOUT THE WRITTEN PERMISSION OF MBARI IS PROHIBITED.

REVISIONS			
REV.	DESCRIPTION	DATE	APPROVED
	INITIAL RELEASE		
A	Modified DFmount and CounterElectrode	9/15/11	hwj

7		1		DuraFET Cover		PEEK (natural)			
Item No.		Qty		Part No.		Description		Material Specification	
Bill of Materials									
Shop / Q.C.		UNLESS OTHERWISE SPECIFIED TOLERANCES ARE:				APPROVALS		DATE	
Inspector		1 PLACE DIM ±1/16		2 PLACE DIM ±.02		3 PLACE DIM ±.005		DRAWN Hans Jannasch 5/30/11	
Shop Foreman		1. DIMENSIONS ARE IN INCHES 2. DIMENSION LIMITS HELD AFTER PLATING. 3. REMOVE BURRS AND SHARP EDGES .015 MAX. 4. MACHINED FILLET RADII .0150-.030. 5. MACHINED DIAMETERS ON A COMMON CENTERLINE CONCENTRIC WITHIN .005 6. PERPENDICULARITY & PARALLELISM OF MACHINED SURFACES .002 PER INCH MAX .012 INCHES FOR SINGLE SURFACE.				ENGINEER		Monterey Bay Aquarium Research Institute MBARI 7700 Sandholdt RD. Moss Landing, CA 95039 PH: (831) 775-1700 	
Engineer						CHECKED			
NEXT ASSY Description		NEXT ASSY Number		PROJECT NO.		900918			
				PROJECT		Novel Sensors			
Application		DO NOT SCALE DRAWING				WEIGHT		SIZE B DWG. NO. 242jahaCommonDF+Ref REV. A	
						SCALE 1:3 CAD FILE: Tornado/chem/SWfiles SHEET 6 OF 6			

Monterey Bay Aquarium Research Institute
MBARI
7700 Sandholdt RD.
Moss Landing, CA 95039
PH: (831) 775-1700

Common DuraFET & Reference