



HOBILabs, Inc.

Instrument Service Document

Instrument: HydroRad-2
Serial Number: HR010424
Institution: MBARI
Contact: Jared Figurski
RMA#: HR2412A

Service Performed:

Incoming inspection and test

Comprehensive calibration (1 channel)

Parts replaced or installed:

Qty	P/N	Description
	---- NONE ----	

Comments:

Instrument arrived with channel B non-functional because of a broken external fiber-optic. The case also was severely corroded. The customer declined repair and requested calibration of channel A. The calibration file returned with the instrument includes a new comprehensive calibration of Channel A, but data from the previous calibration in 2009 for channel B.

NOTE THIS INSTRUMENT IS NOT READY FOR FURTHER SUBMERGED USE WITHOUT ADDITIONAL SERVICE.

Date Completed: 7/14/12

Lead Technician: DRD