

HydroScat-2 Calibration Certification

Cert Date: **January 22, 2008**

Serial Number: **H2010336**

Configuration: **W1B1R0F0S0T112**


Technician: **CE**

Channel 1: bb420

Incremental Response Function

mu:	39.75
mu Rho:	1.1
SigmaExp:	0.143
TempCoeff:	-0.00183

	Gain Factor	Dark Offset
Gain 1:	1.000	12
Gain 2:	9.538	8
Gain 3:	95.006	12
Gain 4:	828.35	10
Gain 5:	9352.7	-8



Channel 2: bb700

Incremental Response Function

mu:	9.536
mu Rho:	1.1
SigmaExp:	0.0145
TempCoeff:	-0.00183

	Gain Factor	Dark Offset
Gain 1:	1.000	12
Gain 2:	9.534	8
Gain 3:	93.305	12
Gain 4:	926.75	8
Gain 5:	9809.7	12

