

HOBİ Labs, Inc.

Instrument Service Document

Instrument: HydroShutter-HS

Serial Number: SS031014

Institution: MBARI

Contact: Erich Rienecker

RMA#: SS1410A

Service Performed:

Internal inspection and functional test

Post-inspection pressure test, 60 operational cycles at 100 PSI

Parts replaced or installed:[illegible]**Comments:**

Date Completed: 3/26/11

Lead Technician: SM