



HOBILabs, Inc.

Instrument Service Document

Instrument: HydroScat-2
Serial Number: H2010337
Institution: MBARI
Contact: Erich Rienecker
RMA#: HS210175F

Service Performed:

Incoming inspection and test
Replace zinc anodes
Calibration
Outgoing inspection and test

Parts replaced or installed:

Qty	P/N	Description
1	MZA-3	Extra-large Zinc Anode
1	MZA-2	Large Zinc Anode

Comments:

Date Completed: 7/31/10

Lead Technician: CE