

Hobi Instrument Services

Service Summary

RMA#: H25021A

Model: HydroScat-2

Completed: March 19, 2022

Serial Number: H2061250

Institution: MBARI

Contact: Jared Figurski

Service Performed:

Incoming inspection and test
Adjust circuit board mounting to fix sensitivity to orientation
Reseat copper faceplate with new PTFE insulation, acrylic screws
Post-servie pressure test and depth calibration
Optical calibration

Comments:

--

[illegible]