



HOBILabs, Inc.

Instrument Service Document

Instrument: HydroScat-2
Serial Number: H2070752
Institution: MBARI
Contact: Jared Figurski
RMA#: H25212A

Service Performed:

Incoming inspection and test

Standard calibration

Outgoing inspection and test

Parts replaced or installed:

Qty	P/N	Description
1	MZA-2	Zinc anode

Comments:

Date Completed: 5/31/12

Lead Technician: SM