



RMA#: H25214A

Completed: π (3/14/15)

Institution: MBARI

Contact: Jared Figurski

[illegible]

Instrument contains an MBARI-installed battery pack which is dead.
--

[illegible]