



---

## Radiometer Service Report

**Date:** 10/17/2023 **SR #:** 1005519551 **Sensor Type:** HyperOCR

**S/N:** 203 **Customer:** MBARI - MONTEREY BAY AQUARIUM RESEARCH INSTITUTE

### Diagnosis:

- No Problems Found.

### Repairs and Modifications:

- Standard Service.
- Passed Calibration.
- The Calibrated output was within 5% of the reference sensor output under natural light conditions.