



Radiometer Service Report

Date: 10/18/2023 **SR #:** 1005519551 **Sensor Type:** HyperOCR I

S/N: 228 **Customer:** MBARI - MONTEREY BAY AQUARIUM RESEARCH INSTITUTE

Diagnosis:

- No Problems Found.

Repairs and Modifications:

- Standard Service.
- Passed Calibration.
- The Calibrated output was within 5% of the reference sensor output under natural light conditions.