



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Monterey Bay Aquarium Research Institute
Job Number 68755
Date 4/11/2012
Technician BLT

<u>Serial Number</u>	052580
----------------------	--------

<u>Low Pressure (PSI)</u>	50	PSI
<u>Time (Minutes)</u>	15	Minutes

<u>High Pressure (PSI)</u>	10000	PSI
<u>Time (Minutes)</u>	30	Minutes

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings.

