

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0385

CALIBRATION DATE: 29-Sep-15

SBE 37 CONDUCTIVITY CALIBRATION DATA

PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.085475e+000

h = 1.559988e-001

i = -4.765889e-004

j = 5.946492e-005

CPcor = -9.5700e-008

CTcor = 3.2500e-006

WBOTC = -7.8271e-005

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2647.30	0.00000	0.00000
1.0000	34.8291	2.97695	5117.77	2.97696	0.00000
4.5000	34.8090	3.28410	5308.00	3.28411	0.00000
15.0000	34.7667	4.26617	5874.75	4.26614	-0.00004
18.4999	34.7573	4.61138	6061.31	4.61139	0.00002
24.0000	34.7476	5.16954	6351.19	5.16956	0.00002
29.0000	34.7424	5.69158	6610.63	5.69156	-0.00001
32.4999	34.7387	6.06398	6789.59	6.06403	0.00006

$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

Conductivity = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$ Siemens / meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = instrument conductivity - bath conductivity

