

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0416
CALIBRATION DATE: 04-Jun-08

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999672e-001

rtca1 = 1.682860e-006

rtca2 = -2.842215e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
0.9999	0.9999690	0.9999689	-0.1
4.4999	0.9999740	0.9999742	0.2
14.9999	0.9999860	0.9999861	0.1
18.4999	0.9999890	0.9999886	-0.4
23.9999	0.9999910	0.9999912	0.2
29.0000	0.9999920	0.9999921	0.1
32.5000	0.9999920	0.9999919	-0.1

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

