



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0416
CALIBRATION DATE: 10-Oct-19

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -2.664609e-004
a1 = 3.304023e-004
a2 = -6.151383e-006
a3 = 2.499102e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	638241.4	1.0000	0.0000
4.5000	547354.5	4.5000	-0.0000
15.0000	351915.0	15.0000	0.0000
18.5000	305593.3	18.5000	-0.0000
24.0000	246244.3	24.0000	0.0000
29.0000	203570.9	29.0000	-0.0000
32.5000	178764.8	32.5000	0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

