

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0417
CALIBRATION DATE: 02-Dec-10

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999692e-001

rtca1 = 1.836409e-006

rtca2 = -2.910280e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
0.9999	0.9999710	0.9999710	0.0
4.4999	0.9999770	0.9999769	-0.1
15.0000	0.9999900	0.9999902	0.2
18.5000	0.9999930	0.9999932	0.2
24.0000	0.9999970	0.9999965	-0.5
29.0000	0.9999980	0.9999980	-0.0
32.5000	0.9999980	0.9999982	0.2

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

