

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0470

CALIBRATION DATE: 21-Oct-15

SBE 37 RTC CALIBRATION DATA

ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999754e-001

rtca1 = 1.803577e-006

rtca2 = -3.043588e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
1.0000	0.9999770	0.9999772	0.2
4.5000	0.9999830	0.9999829	-0.1
15.0000	0.9999960	0.9999956	-0.4
18.5000	0.9999980	0.9999983	0.3
24.0000	1.0000010	1.0000011	0.1
29.0000	1.0000020	1.0000021	0.1
32.5000	1.0000020	1.0000018	-0.2

RTC frequency = $rtca0 + rtca1 * t + rtca2 * t^2$

Residual = (Computed RTC frequency - Measured RTC frequency) * 1e6

