

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0471
CALIBRATION DATE: 06-Nov-08

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999729e-001

rtca1 = 1.918790e-006

rtca2 = -2.726698e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
1.0000	0.9999750	0.9999748	-0.2
4.4999	0.9999810	0.9999810	0.0
15.0000	0.9999950	0.9999956	0.6
18.4999	0.9999990	0.9999991	0.1
23.9999	1.0000040	1.0000033	-0.7
28.9999	1.0000060	1.0000057	-0.3
32.5001	1.0000060	1.0000065	0.5

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

