

# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0471  
CALIBRATION DATE: 07-Jul-10

SBE 37 RTC CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

## COEFFICIENTS:

rtca0 = 9.999729e-001

rtca1 = 1.918770e-006

rtca2 = -2.726641e-008

| BATH TEMP<br>(ITS-90) | RTC FREQ<br>(Hz) | COMPUTED FREQ<br>(Hz) | RESIDUAL<br>(PPM) |
|-----------------------|------------------|-----------------------|-------------------|
| 0.9999                | 0.9999750        | 0.9999748             | -0.2              |
| 4.5000                | 0.9999810        | 0.9999810             | 0.0               |
| 15.0000               | 0.9999950        | 0.9999956             | 0.6               |
| 18.5000               | 0.9999990        | 0.9999991             | 0.1               |
| 24.0000               | 1.0000040        | 1.0000033             | -0.7              |
| 29.0000               | 1.0000060        | 1.0000057             | -0.3              |
| 32.5000               | 1.0000060        | 1.0000065             | 0.5               |

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

