



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0473
CALIBRATION DATE: 20-Nov-20

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -5.427783e-005
a1 = 2.827497e-004
a2 = -3.018335e-006
a3 = 1.682112e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	783581.0	1.0001	0.0001
4.5000	669841.8	4.4999	-0.0001
15.0000	426626.9	15.0001	0.0001
18.5000	369334.9	18.5000	-0.0000
24.0000	296187.4	24.0000	-0.0000
29.0000	243807.4	28.9999	-0.0001
32.5000	213457.5	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

