



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0475
CALIBRATION DATE: 13-Dec-17

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 6.608137e-006
a1 = 2.747622e-004
a2 = -2.389394e-006
a3 = 1.535493e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	703635.3	1.0001	0.0001
4.5000	601551.5	4.4999	-0.0001
15.0000	383231.5	15.0000	0.0000
18.5000	331790.8	18.5001	0.0001
24.0000	266112.5	24.0000	-0.0000
29.0000	219073.9	28.9998	-0.0002
32.5000	191814.1	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

