

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0661
CALIBRATION DATE: 02-Jul-10

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

a0 = -1.038055e-004
a1 = 2.943941e-004
a2 = -3.692431e-006
a3 = 1.894702e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	691065.8	1.0000	0.0000
4.5000	592028.6	4.4999	-0.0001
15.0000	379410.9	15.0001	0.0001
18.5000	329110.4	18.5000	0.0000
24.0000	264733.6	23.9999	-0.0001
29.0000	218502.1	29.0000	-0.0000
32.5000	191655.2	32.5000	0.0000

Temperature ITS-90 = $1/[a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]] - 273.15$ (°C)

Residual = instrument temperature - bath temperature

