

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0811
CALIBRATION DATE: 25-Apr-12

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.080139e+000
h = 1.503203e-001
i = -3.795766e-004
j = 5.134967e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -7.3368e-005

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2688.57	0.00000	0.00000
1.0000	34.9210	2.98406	5210.02	2.98405	-0.00001
4.5000	34.9009	3.29192	5403.90	3.29193	0.00001
15.0000	34.8583	4.27622	5981.38	4.27622	-0.00000
18.5000	34.8493	4.62227	6171.45	4.62228	0.00001
24.0000	34.8397	5.18172	6466.75	5.18172	-0.00000
29.0000	34.8348	5.70501	6731.06	5.70499	-0.00001
32.5000	34.8327	6.07853	6913.45	6.07854	0.00001

$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$

Date, Slope Correction

