

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0812
CALIBRATION DATE: 08-Dec-09

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999752e-001

rtca1 = 1.675645e-006

rtca2 = -2.609446e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
1.0001	0.9999770	0.9999768	-0.2
4.5000	0.9999820	0.9999822	0.2
14.9999	0.9999940	0.9999944	0.4
18.5000	0.9999980	0.9999973	-0.7
24.0000	1.0000000	1.0000004	0.4
29.0000	1.0000020	1.0000018	-0.2
32.4999	1.0000020	1.0000021	0.1

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

