

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0814
 CALIBRATION DATE: 11-Jun-09

SBE 37 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.071959e+000	CPcor = -9.5700e-008
h = 1.541963e-001	CTcor = 3.2500e-006
i = -4.266562e-004	WBOTC = -7.7537e-005
j = 5.868642e-005	

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2645.06	0.00000	0.00000
1.0000	34.6680	2.96449	5127.33	2.96451	0.00002
4.5000	34.6486	3.27046	5318.22	3.27045	-0.00001
15.0000	34.6065	4.24859	5886.76	4.24856	-0.00003
18.5000	34.5974	4.59245	6073.87	4.59245	-0.00001
24.0000	34.5872	5.14830	6364.53	5.14833	0.00003
29.0000	34.5817	5.66820	6624.64	5.66823	0.00002
32.4999	34.5791	6.03928	6804.07	6.03925	-0.00003

$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$

