

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0816
 CALIBRATION DATE: 07-Mar-07

SBE 37 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.061390e+000	CPcor = -9.5700e-008
h = 1.475521e-001	CTcor = 3.2500e-006
i = -2.686046e-004	WBOTC = -8.0370e-005
j = 4.230356e-005	

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2688.21	0.00000	0.00000
1.0000	34.8912	2.98176	5239.23	2.98176	0.00000
4.5000	34.8724	3.28949	5435.09	3.28949	-0.00000
15.0000	34.8311	4.27324	6018.32	4.27324	0.00000
18.5000	34.8225	4.61910	6210.26	4.61911	0.00001
24.0000	34.8130	5.17819	6508.42	5.17816	-0.00002
29.0000	34.8079	5.70110	6775.31	5.70112	0.00002
32.5000	34.8048	6.07421	6959.37	6.07420	-0.00001

$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$

