

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0816
CALIBRATION DATE: 12-Jun-09

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999795e-001

rtca1 = 1.697299e-006

rtca2 = -2.911458e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
0.9999	0.9999810	0.9999812	0.2
4.4999	0.9999870	0.9999866	-0.4
15.0001	0.9999980	0.9999985	0.5
18.5000	1.0000010	1.0000010	-0.0
24.0001	1.0000040	1.0000035	-0.5
29.0000	1.0000040	1.0000043	0.3
32.5000	1.0000040	1.0000040	-0.0

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

