

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0817
CALIBRATION DATE: 08-Dec-16

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -2.082669e-004
a1 = 3.154750e-004
a2 = -5.085676e-006
a3 = 2.204543e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	685476.9	1.0000	-0.0000
4.5000	587353.3	4.5000	0.0000
15.0000	376694.7	15.0000	-0.0000
18.5000	326849.0	18.5000	0.0000
24.0000	263048.3	24.0000	0.0000
29.0000	217224.8	29.0000	-0.0000
32.5000	190610.0	32.5000	0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1/\{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

