

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0818
CALIBRATION DATE: 12-Jun-09

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.069660e+000
h = 1.548525e-001
i = -3.801131e-004
j = 5.338389e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -7.3731e-005

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2635.74	0.00000	0.00000
0.9999	34.5767	2.95742	5108.81	2.95736	-0.00006
4.4999	34.5537	3.26237	5298.94	3.26245	0.00008
15.0001	34.5082	4.23781	5865.25	4.23782	0.00001
18.5000	34.4972	4.58058	6051.55	4.58056	-0.00003
24.0001	34.4853	5.13481	6341.09	5.13480	-0.00002
29.0000	34.4753	5.65272	6600.02	5.65273	0.00002

$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p)$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Residual = instrument conductivity - bath conductivity

