

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0818
CALIBRATION DATE: 30-Oct-13

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.070137e+000
h = 1.550992e-001
i = -4.494301e-004
j = 5.876991e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -7.3731e-005

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2635.46	0.00000	0.00000
1.0000	34.6908	2.96626	5114.20	2.96627	0.00001
4.4999	34.6708	3.27234	5304.73	3.27233	-0.00000
15.0000	34.6279	4.25094	5872.24	4.25093	-0.00001
18.5000	34.6190	4.59501	6059.02	4.59500	-0.00001
24.0000	34.6093	5.15123	6349.20	5.15125	0.00002
29.0000	34.6046	5.67153	6608.91	5.67154	0.00000
32.5000	34.6022	6.04286	6788.07	6.04286	-0.00001

$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$

Date, Slope Correction

