

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0819
CALIBRATION DATE: 17-Apr-12

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.097973e+000
h = 1.567950e-001
i = -6.517533e-004
j = 7.427055e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -7.4742e-005

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2658.68	0.00000	0.00000
1.0000	34.9254	2.98440	5125.58	2.98446	0.00006
4.5000	34.9048	3.29225	5315.56	3.29217	-0.00007
15.0000	34.8606	4.27647	5881.82	4.27648	0.00001
18.5000	34.8510	4.62247	6068.17	4.62247	-0.00000
24.0000	34.8405	5.18183	6357.70	5.18183	0.00001
29.0000	34.8344	5.70495	6616.80	5.70497	0.00002
32.5000	34.8308	6.07823	6795.49	6.07821	-0.00002

$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$

