



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0823
CALIBRATION DATE: 19-Nov-23

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999672e-001
rtca1 = 1.843667e-006
rtca2 = -3.143147e-008

BATH TEMPERATURE (° C)	RTC FREQUENCY (Hz)	COMPUTED FREQUENCY (Hz)	RESIDUAL (PPM)
1.0000	0.9999690	0.9999691	0.1
4.4999	0.9999750	0.9999749	-0.1
14.9999	0.9999880	0.9999878	-0.2
18.5000	0.9999900	0.9999906	0.6
24.0000	0.9999940	0.9999934	-0.6
29.0000	0.9999940	0.9999943	0.3
32.5000	0.9999940	0.9999940	-0.0

t = temperature (°C)

RTC frequency = $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) * 1e6

