

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0825
CALIBRATION DATE: 17-Apr-12

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999756e-001

rtca1 = 1.706741e-006

rtca2 = -2.761592e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
1.0000	0.9999770	0.9999773	0.3
4.5000	0.9999830	0.9999827	-0.3
15.0000	0.9999950	0.9999950	-0.0
18.5000	0.9999980	0.9999977	-0.3
24.0000	1.0000000	1.0000006	0.6
29.0000	1.0000020	1.0000019	-0.1
32.5000	1.0000020	1.0000019	-0.1

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

