



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0825
CALIBRATION DATE: 08-Sep-21

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.197243e-001
h = 1.413135e-001
i = -1.114485e-004
j = 3.726446e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -2.9023e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2552.35	0.00000	0.00000
1.0000	34.6690	2.96457	5234.81	2.96460	0.00002
4.4999	34.6496	3.27053	5436.23	3.27050	-0.00004
15.0000	34.6075	4.24870	6034.51	4.24870	-0.00000
18.4999	34.5988	4.59261	6230.93	4.59265	0.00004
23.9999	34.5905	5.14873	6535.67	5.14869	-0.00004
29.0000	34.5852	5.66871	6807.99	5.66872	0.00001
32.5001	34.5800	6.03944	6995.57	6.03969	0.00025

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

